

# **Ophthalmic Microsurgical Knives**

# With or Without Polycarbonate Handle

- Implant Tip Size 4.0, 5.0, 5.1, 5.2, 5.5 mm
- Phaco Slit Tip Size 1.0, 1.2, 1.4, 1.5, 1.65, 1.8, 2.0, 2.2, 2.4, 2.5, 2.6, 2.65, 2.75, 2.8, 3.0, 3.2, 3.5, 5.2 mm
- Crescent Bevelled Up, Bevelled Down, Double Bevelled Up, Double Bevelled Down.
- Stab Incision 15/22/30/45 degree with restricted depth 3.0 mm to 6.0 mm in 15 degree and restricted depth 3.0 mm in 30 degree.
- MVR Angled and Straight in 19G, 20G, 21G, 22G, 23G & 24G
- Customized (Size and Style)

## **Indications**

The device is indicated to be used for small Incision Cataract Surgical Procedures such as ECCE, Manual Phaco and Ultra Sonic Phaco Surgery of the Eye. The device is also used for corneal / scleral incisions & cutting sutures in ophthalmic surgical procedures.

# Instructions for Use:

- Pick up one pouch and open from peel to open side. Remove pouch and pick up tray.

- Remove knife from tray without touching blade tip & edges.
  The product should be used immediately once the packing is opened.
- Dispose it after use. It should be disposed off properly as per country's regulations and disposal of medical bio waste rules.
- The user must be fully acquainted with surgical procedures and techniques. The product should be handled according to the norms.

- Warning: Read instruction before use.
- Select an appropriate type of knife for each surgical purpose. Do not use for other purpose than ophthalmic surgery.
- Do not use a knife if it has been hit against or in contact with other parts than the eyeball.
- Do not reuse or re-sterilize. Reuse or re-sterilization could compromise seriously the structural integrity and/or might lead to damage or infection to
- If it is reused, it may cause serious problems to patients such as redness in the eye, blindness and other communicable diseases
- One must inspect and check the product before use. Should not be used if the packing is damaged.



# **Ophthalmic Microsurgical Knives**

# With or Without Polycarbonate Handle

- Implant Tip Size 4.0, 5.0, 5.1, 5.2, 5.5 mm
- Phaco Slit Tip Size 1.0, 1.2, 1.4, 1.5, 1.65, 1.8, 2.0, 2.2, 2.4, 2.5, 2.6, 2.65, 2.75, 2.8, 3.0, 3.2, 3.5, 5.2 mm
- Crescent Bevelled Up, Bevelled Down, Double Bevelled Up, Double Bevelled Down.
- Stab Incision 15/22/30/45 degree with restricted depth 3.0 mm to 6.0 mm in 15 degree and restricted depth 3.0 mm in 30 degree.
- MVR Angled and Straight in 19G, 20G, 21G, 22G, 23G & 24G
- Customized (Size and Style)

# **Indications**

The device is indicated to be used for small Incision Cataract Surgical Procedures such as ECCE, Manual Phaco and Ultra Sonic Phaco Surgery of the Eye. The device is also used for corneal / scleral incisions & cutting sutures in ophthalmic surgical procedures.

# Instructions for Use:

- Pick up one pouch and open from peel to open side. Remove pouch and pick up tray.

- Remove knife from tray without touching blade tip & edges.
  The product should be used immediately once the packing is opened.
  Dispose it after use. It should be disposed off properly as per country's
- regulations and disposal of medical bio waste rules.
- The user must be fully acquainted with surgical procedures and techniques. The product should be handled according to the norms.

# Warning:

- Read instruction before use.
- Select an appropriate type of knife for each surgical purpose. Do not use for other purpose than ophthalmic surgery.
- Do not use a knife if it has been hit against or in contact with other parts than the eyeball.
- Do not reuse or re-sterilize. Reuse or re-sterilization could compromise seriously the structural integrity and/or might lead to damage or infection to
- If it is reused, it may cause serious problems to patients such as redness in the eye, blindness and other communicable diseases.
- One must inspect and check the product before use. Should not be used if the packing is damaged.



# Ophthalmic Microsurgical Knives

# With or Without Polycarbonate Handle

- Implant Tip Size 4.0, 5.0, 5.1, 5.2, 5.5 mm
- Phaco Slit Tip Size 1.0, 1.2, 1.4, 1.5, 1.65, 1.8, 2.0, 2.2, 2.4, 2.5, 2.6, 2.65, 2.75, 2.8, 3.0, 3.2, 3.5, 5.2 mm
- Crescent Bevelled Up, Bevelled Down, Double Bevelled Up, Double Bevelled Down.
- Stab Incision 15/22/30/45 degree with restricted depth 3.0 mm to 6.0 mm in 15 degree and restricted depth 3.0 mm in 30 degree.
- MVR Angled and Straight in 19G, 20G, 21G, 22G, 23G & 24G
- Customized (Size and Style)

### Indications

The device is indicated to be used for small Incision Cataract Surgical Procedures such as ECCE, Manual Phaco and Ultra Sonic Phaco Surgery of the Eye. The device is also used for corneal / scleral incisions & cutting sutures in ophthalmic surgical procedures.

### Instructions for Use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray. Remove knife from tray without touching blade tip & edges.
- The product should be used immediately once the packing is opened. Dispose it after use. It should be disposed off properly as per country's
- regulations and disposal of medical bio waste rules.
  The user must be fully acquainted with surgical procedures and techniques.
- The product should be handled according to the norms.

- Warning:
  Read instruction before use.
- Select an appropriate type of knife for each surgical purpose.
- Do not use for other purpose than ophthalmic surgery.

  Do not use a knife if it has been hit against or in contact with other parts than
- Do not reuse or re-sterilize. Reuse or re-sterilization could compromise seriously the structural integrity and/or might lead to damage or infection to patients.
- If it is reused, it may cause serious problems to patients such as redness in the eve, blindness and other communicable diseases.
- One must inspect and check the product before use.
- Should not be used if the packing is damaged.



# **Ophthalmic Microsurgical Knives**

# With or Without Polycarbonate Handle

- Implant Tip Size 4.0, 5.0, 5.1, 5.2, 5.5 mm
- Phaco Slit Tip Size 1.0, 1.2, 1.4, 1.5, 1.65, 1.8, 2.0, 2.2, 2.4, 2.5, 2.6, 2.65, 2.75, 2.8, 3.0, 3.2, 3.5, 5.2 mm
- Crescent Bevelled Up, Bevelled Down, Double Bevelled Up, Double Bevelled Down.
- Stab Incision 15/22/30/45 degree with restricted depth 3.0 mm to 6.0 mm in 15 degree and restricted depth 3.0 mm in 30 degree.
- MVR Angled and Straight in 19G, 20G, 21G, 22G, 23G & 24G
- Customized (Size and Style)

The device is indicated to be used for small Incision Cataract Surgical Procedures such as ECCE, Manual Phaco and Ultra Sonic Phaco Surgery of the Eye. The device is also used for corneal / scleral incisions & cutting sutures in ophthalmic surgical procedures.

# Instructions for Use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray. Remove knife from tray without touching blade tip & edges.
- The product should be used immediately once the packing is opened. Dispose it after use. It should be disposed off properly as per country's regulations and disposal of medical bio waste rules.
- The user must be fully acquainted with surgical procedures and techniques. The product should be handled according to the norms.

# Warning:

- Read instruction before use.
- Select an appropriate type of knife for each surgical purpose.
- Do not use for other purpose than ophthalmic surgery.

  Do not use a knife if it has been hit against or in contact with other parts than the eyeball.

  Do not reuse or re-sterilize. Reuse or re-sterilization could compromise
- seriously the structural integrity and/or might lead to damage or infection to patients. If it is reused, it may cause serious problems to patients such as redness in
- the eve. blindness and other communicable diseases. One must inspect and check the product before use.
- Should not be used if the packing is damaged.

- The product is non-toxic, sterile and non-pyrogenic if the packaging is intact. Ensure safe and proper disposal as per law of the land for handling of Bio Medical Waste used / discarded product and packaging to avoid adverse effect to environment, children and stray animals.
- The device is for intraocular use only. Do not use the product after expiry date.

Store at a temperature between 10°C to 40°C and a relative humidity less than 65%. Do not expose to direct sunlight. Expiry date on the pack indicates the period within which the device should be used.

### Contraindications:

If it is used for any other purposes than indicated use, will be the responsibility

If product is used by uncompetent / untrained personnel.

## Explication of Symbol:

Symbol	Meaning
$\triangle$	Caution
ш	Manufacturer
\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\	Date and Country of Manufacture
Ţ <u>i</u>	Consult instructions for use
类	Keep away from sunlight
MD	Medical Device
	Do not use hook
<u>11</u>	This side up
10°C 40°C	Temperature Limit
35%	Humidity Limitation
<b>(fi</b> )	Keep Out of reach of children
N N N N N N N N N N N N N N N N N N N	Handle with care

Symbol	Meaning
8	Maximum Stacking
LOT	Batch Code
Ω	Use by date
<b>®</b>	Do not use if package is damaged
2	Do not re-use
	Peel Open
<b>†</b>	Keep DRY
STERILEEO	Single Sterile Barrier System, sterilized using Ethylene Oxide
STEMLE EO	Single Sterile Barrier System with protective packaging outside, sterilized using Ethylene Oxide
<b>3</b>	Do not re-sterilize
	Distributor

TARUN ENTERPRISES
Works. Shed No. E20 NSIC Campus, Mirzapur Road
Naini, Prayagraj-211009, UP, India
Email: info@optitecheyecare.com, 19+91-8176080204 Regd. Off. 8/8, Strachy Road Prayagraj-211001, UP, India MFG. LIC.NO.:-MFG/MD/2023/000431

REF TE/ART/2015-16/01 L. Rev. No. 08 dated 10th May 2023

- The product is non-toxic, sterile and non-pyrogenic if the packaging is intact.
- Ensure safe and proper disposal as per law of the land for handling of Bio Medical Waste used / discarded product and packaging to avoid adverse effect to environment, children and stray animals. The device is for intraocular use only.
- Do not use the product after expiry date.

# Storage:

Store at a temperature between 10°C to 40°C and a relative humidity less than 65%. Do not expose to direct sunlight. Expiry date on the pack indicates the period within which the device should be used.

# Contraindications:

If it is used for any other purposes than indicated use, will be the responsibility

If product is used by uncompetent / untrained personnel.

# **Explication of Symbol:**

Symbol	Meaning
$\triangle$	Caution
ш	Manufacturer
W.	Date and Country of Manufacture
[]i	Consult instructions for use
类	Keep away from sunlight
MD	Medical Device
8	Do not use hook
11	This side up
10°C	Temperature Limit
35%	Humidity Limitation
<b>(fi)</b>	Keep Out of reach of children
*	Handle with care

Symbol	Meaning
8	Maximum Stacking
LOT	Batch Code
Ω	Use by date
8	Do not use if package is damaged
2	Do not re-use
	Peel Open
<b>T</b>	Keep DRY
STERLED	Single Sterile Barrier System, sterilized using Ethylene Oxide
(STEPILE EO)	Single Sterile Barrier System with protective packaging outside, sterilized using Ethylene Oxide
3	Do not re-sterilize
	Distributor

TARUN ENTERPRISES
Works. Shed No. E20 NSIC Campus, Mirzapur Road
Naini, Prayagraj-211009, UP, India
Email: info@optitecheyecare.com, ᠖+91-8176080204 Regd. Off. 8/8, Strachy Road Prayagraj-211001, UP, India

MFG. LIC.NO.:-MFG/MD/2023/000431

REF TE/ART/2015-16/01 L. Rev. No. 08 dated 10th May 2023

- The product is non-toxic, sterile and non-pyrogenic if the packaging is intact. Ensure safe and proper disposal as per law of the land for handling of Bio Medical Waste used / discarded product and packaging to avoid adverse effect to environment, children and stray animals.
- The device is for intraocular use only.

  Do not use the product after expiry date.

Store at a temperature between 10°C to 40°C and a relative humidity less than 65%. Do not expose to direct sunlight. Expiry date on the pack indicates the period within which the device should be used.

### Contraindications:

If it is used for any other purposes than indicated use, will be the responsibility If product is used by uncompetent / untrained personnel.

# **Explication of Symbol:**

Symbol	Meaning
$\triangle$	Caution
	Manufacturer
\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\	Date and Country of Manufacture
[]i	Consult instructions for use
类	Keep away from sunlight
MD	Medical Device
8	Do not use hook
<u>11</u>	This side up
10°C	Temperature Limit
35%	Humidity Limitation
<b>(fi)</b>	Keep Out of reach of children
8	Handle with care

Symbol	Meaning
8	Maximum Stacking
LOT	Batch Code
₽	Use by date
<b>®</b>	Do not use if package is damaged
2	Do not re-use
	Peel Open
<b>†</b>	Keep DRY
STERLEO	Single Sterile Barrier System, sterilized using Ethylene Oxide
STERLEGO	Single Sterile Barrier System with protective packaging outside, sterilized using Ethylene Oxide
3	Do not re-sterilize
	Distributor

TARUN ENTERPRISES
Works. Shed No. E20 NSIC Campus, Mirzapur Road
Naini, Prayagraj-211009, UP, India
Emaii: info@optitecheyecare.com, lie+91-8176080204
Regd. Off. 8/8,Strachy Road Prayagraj-211001,UP,India

MFG. LIC.NO.:-MFG/MD/2023/000431

REF TE/ART/2015-16/01 L. Rev. No. 08 dated 10th May 2023

- The product is non-toxic, sterile and non-pyrogenic if the packaging is intact. Ensure safe and proper disposal as per law of the land for handling of Bio
- Medical Waste used / discarded product and packaging to avoid adverse effect to environment, children and stray animals.
- The device is for intraocular use only.

  Do not use the product after expiry date.

Store at a temperature between 10°C to 40°C and a relative humidity less than 65%. Do not expose to direct sunlight. Expiry date on the pack indicates the period within which the device should be used.

# Contraindications:

If it is used for any other purposes than indicated use, will be the responsibility

If product is used by uncompetent / untrained personnel.

# **Explication of Symbol:**

•	-
Symbol	Meaning
$\triangle$	Caution
	Manufacturer
CW.	Date and Country of Manufacture
[]i	Consult instructions for use
类	Keep away from sunlight
MD	Medical Device
<b>8</b>	Do not use hook
<u>11</u>	This side up
10°C	Temperature Limit
35%	Humidity Limitation
(fr)	Keep Out of reach of children
•	Handle with care

Symbol	Meaning
8	Maximum Stacking
LOT	Batch Code
Ω	Use by date
8	Do not use if package is damaged
8	Do not re-use
	Peel Open
<b>†</b>	Keep DRY
STERLED	Single Sterile Barrier System, sterilized using Ethylene Oxide
(STEPILE EO)	Single Sterile Barrier System with protective packaging outside, sterilized using Ethylene Oxide
8	Do not re-sterilize
	Distributor

TARUN ENTERPRISES
Works. Shed No. E20 NSIC Campus, Mirzapur Road
Naini, Prayagraj-211009, UP, India
Email: info@optitecheyecare.com, 1891-8176080204
Regd. Off. 8/8,Strachy Road Prayagraj-211001,UP,India

MFG. LIC.NO.:-MFG/MD/2023/000431 REF TE/ART/2015-16/01 L. Rev. No. 08 dated 10th May 2023