

Ophthalmic Surgical Instruments -
OPHTHALMIC CANNULA

Indication:

The surgical instruments are indicated to use to hold soft tissues during surgery, to tie sutures and for cutting tissues

Instruction for use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray.
- Dispose it after use, it should be disposed off properly as per regulatory law of country for the disposal of medical bio waste
- Store it in a cool and dry place.
- Do not expose to heat or direct sunlight.
- Device is to be used by qualified doctors/competent persons.
- The product should be handled according to the norms.
- These are EO sterile instrument

Warning


- Read instruction before use.
- The products should be Sterilized/Autoclaved before re-use.
- Should not be used if the packing is damaged.
- If product is resterilized or reused it may have impact on mechanical properties and intended use of the product.

Storage:

Store at a temperature between 10°C to 40°C and a relative humidity in between 35% to 65 %. Do not expose to direct sunlight. Expiry date on that pack indicates the period within which the device should be used.

Contraindications:

- If it is used for any other purposes than indicated use, will be the responsibility of the user.
- If product is used by incompetent/untrained personnel



TARUN ENTERPRISES

Works : Shed No. E 20 NSIC Campus, Mirzapur Road
Naini, Prayagraj (Allahabad) - 211009, U.P., INDIA
Regd. Off.: 8/8, Strachy Road
Prayagraj (Allahabad) - 211001, U.P., INDIA
e-mail: info@optitecheyecare.com
+91 8176080204

Mfg. Lic. No.:MFG/MD/2023/000431

REF/TE/ART/2015-16/01 L Rev. No. 07 Dated 10th Sep 2024

Ophthalmic Surgical Instruments -
OPHTHALMIC CANNULA

Indication:

The surgical instruments are indicated to use to hold soft tissues during surgery, to tie sutures and for cutting tissues

Instruction for use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray.
- Dispose it after use, it should be disposed off properly as per regulatory law of country for the disposal of medical bio waste
- Store it in a cool and dry place.
- Do not expose to heat or direct sunlight.
- Device is to be used by qualified doctors/competent persons.
- The product should be handled according to the norms.
- These are EO sterile instrument

Warning

- Read instruction before use.
- The products should be Sterilized/Autoclaved before re-use.
- Should not be used if the packing is damaged.
- If product is resterilized or reused it may have impact on mechanical properties and intended use of the product.

Storage:

Store at a temperature between 10°C to 40°C and a relative humidity in between 35% to 65 %. Do not expose to direct sunlight. Expiry date on that pack indicates the period within which the device should be used.

Contraindications:

- If it is used for any other purposes than indicated use, will be the responsibility of the user.
- If product is used by incompetent/untrained personnel



TARUN ENTERPRISES

Works : Shed No. E 20 NSIC Campus, Mirzapur Road
Naini, Prayagraj (Allahabad) - 211009, U.P., INDIA
Regd. Off.: 8/8, Strachy Road
Prayagraj (Allahabad) - 211001, U.P., INDIA
e-mail: info@optitecheyecare.com
+91 8176080204

Mfg. Lic. No.:MFG/MD/2023/000431

REF/TE/ART/2015-16/01 L Rev. No. 07 Dated 10th Sep 2024

Ophthalmic Surgical Instruments -
OPHTHALMIC CANNULA

Indication:

The surgical instruments are indicated to use to hold soft tissues during surgery, to tie sutures and for cutting tissues

Instruction for use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray.
- Dispose it after use, it should be disposed off properly as per regulatory law of country for the disposal of medical bio waste
- Store it in a cool and dry place.
- Do not expose to heat or direct sunlight.
- Device is to be used by qualified doctors/competent persons.
- The product should be handled according to the norms.
- These are EO sterile instrument

Warning


- Read instruction before use.
- The products should be Sterilized/Autoclaved before re-use.
- Should not be used if the packing is damaged.
- If product is resterilized or reused it may have impact on mechanical properties and intended use of the product.

Storage:

Store at a temperature between 10°C to 40°C and a relative humidity in between 35% to 65 %. Do not expose to direct sunlight. Expiry date on that pack indicates the period within which the device should be used.

Contraindications:

- If it is used for any other purposes than indicated use, will be the responsibility of the user.
- If product is used by incompetent/untrained personnel



TARUN ENTERPRISES

Works : Shed No. E 20 NSIC Campus, Mirzapur Road
Naini, Prayagraj (Allahabad) - 211009, U.P., INDIA
Regd. Off.: 8/8, Strachy Road
Prayagraj (Allahabad) - 211001, U.P., INDIA
e-mail: info@optitecheyecare.com
+91 8176080204

Mfg. Lic. No.:MFG/MD/2023/000431

REF/TE/ART/2015-16/01 L Rev. No. 07 Dated 10th Sep 2024

Ophthalmic Surgical Instruments -
OPHTHALMIC CANNULA

Indication:

The surgical instruments are indicated to use to hold soft tissues during surgery, to tie sutures and for cutting tissues

Instruction for use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray.
- Dispose it after use, it should be disposed off properly as per regulatory law of country for the disposal of medical bio waste
- Store it in a cool and dry place.
- Do not expose to heat or direct sunlight.
- Device is to be used by qualified doctors/competent persons.
- The product should be handled according to the norms.
- These are EO sterile instrument

Warning

- Read instruction before use.
- The products should be Sterilized/Autoclaved before re-use.
- Should not be used if the packing is damaged.
- If product is resterilized or reused it may have impact on mechanical properties and intended use of the product.

Storage:

Store at a temperature between 10°C to 40°C and a relative humidity in between 35% to 65 %. Do not expose to direct sunlight. Expiry date on that pack indicates the period within which the device should be used.

Contraindications:

- If it is used for any other purposes than indicated use, will be the responsibility of the user.
- If product is used by incompetent/untrained personnel



TARUN ENTERPRISES

Works : Shed No. E 20 NSIC Campus, Mirzapur Road
Naini, Prayagraj (Allahabad) - 211009, U.P., INDIA
Regd. Off.: 8/8, Strachy Road
Prayagraj (Allahabad) - 211001, U.P., INDIA
e-mail: info@optitecheyecare.com
+91 8176080204

Mfg. Lic. No.:MFG/MD/2023/000431

REF/TE/ART/2015-16/01 L Rev. No. 07 Dated 10th Sep 2024

Ophthalmic Surgical Instruments - **OPHTHALMIC CANNULA**

Indication:

The surgical instruments are indicated to use to hold soft tissues during surgery, to tie sutures and for cutting tissues

Instruction for use:

- Pick up one pouch and open from peel to open side.
- Remove pouch and pick up tray.
- Dispose it after use, it should be disposed off properly as per regulatory law of country for the disposal of medical bio waste
- Store it in a cool and dry place.
- Do not expose to heat or direct sunlight.
- Device is to be used by qualified doctors/competent persons.
- The product should be handled according to the norms.
- These are EO sterile instrument

Warning


- Read instruction before use.
- The products should be Sterilized/Autoclaved before re-use.
- Should not be used if the packing is damaged.
- If product is resterilized or reused it may have impact on mechanical properties and intended use of the product.

Storage:

Store at a temperature between 10 °C to 40 °C and a relative humidity in between 35 RH to 65 RH. Do not expose to direct sunlight. Expiry date on that pack indicates the period within which the device should be used.

Contraindications:

- If it is used for any other purposes than indicated use, will be the responsibility of the user.
- If product is used by incompetent/untrained personnel



TARUN ENTERPRISES

Works : Shed No. E 20 NSIC Campus, Mirzapur Road
Naini, Prayagraj (Allahabad) - 211009, U.P., INDIA

Regd. Off.: 8/8, Strachy Road
Prayagraj (Allahabad) - 211001, U.P., INDIA

e-mail: info@optitecheyecare.com

 +91 8176080204